

Catalog Number AS 015 BG

## Highlights:

- Results in 5 minutes or less
- Available as 100-strip kits, in bulk packaging, or in QuickCombs™

## Contents of Kit:

- 100 QuickStix Strips packed in two moisture-resistant canisters
- 100 transfer pipettes
- 100 reaction vials

## Items Not Provided:

- Waring blender, model 31BL91 or equivalent
- Glass jar adapter (Eberbach #E8495)
- Glass Mason jars
- Graduated cylinder
- Tap water
- Protective cover for blender jar while grinding

## Intended Use

This EnviroLogix QuickStix™ Kit for YieldGard® Rootworm Bulk Grain is designed to extract and detect Cry3Bb protein at the levels typically expressed in YieldGard Rootworm corn. The sensitivity of these QuickStix strips is 0.5% (i.e. 1 kernel in a pool of 200).

## How the Test Works

In order to detect the Cry3Bb protein expressed by YieldGard Rootworm corn, the sample must first be extracted in tap water to solubilize the protein. Each QuickStix strip has an absorbent pad at each end. The protective tape with the arrow indicates the end of the strip to insert into the reaction vial. The sample will travel up the membrane strip and be absorbed into the larger pad at the top of the strip. The portion of the strip between the protective tape and the absorbent pad at the top of the strip is used to view the reactions as described under “Interpreting the Results”. Please avoid bending the strips.

## Sample Preparation

### Step 1: Determine Number and Size of Sub-samples

1. Collect a composite sample according to USDA/ GIPSA instructions found in the following reference documents:
  - <http://www.archive.gipsa.usda.gov/reference-library/handbooks/grain-insp/grbook1/bk1.pdf> - USDA Grain Inspection Handbook, Book 1, Grain Sampling.
  - <http://www.archive.gipsa.usda.gov/biotech/sample2.htm> - Guidance document entitled Sampling for the Detection of Biotech Grains.
  - <http://www.archive.gipsa.usda.gov/biotech/sample1.htm> - Practical Application of Sampling for the Detection of Biotech Grains.
  - <http://www.archive.gipsa.usda.gov/biotech/samplingplan1.xls> - This website provides a simple to use Sample Planner (29K Excel Spreadsheet). The planner allows you to enter different assumptions in terms of sample size, number of samples, acceptable quality level and to determine the probability of accepting lots with given concentration levels. It also plots the probabilities in graph form for easy interpretation. Specific data can be saved for documentation and future analyses.
2. The following is a helpful reference for use in designing a sampling plan: Remund, K.M., Dixon, D.A., Wright D.L., Holden, L.R. “Statistical considerations in seed purity testing for transgenic traits”, Seed Science Research, June 2001, Vol. 11 No.2, pp. 101-119.
3. To select the appropriate sample size, determine the purity standard and the degree of confidence required. Confidence level means the statistical probability that the % of YieldGard Rootworm corn in the lot is below the selected purity standard. Table 1 provides a guideline for determining the number of sub-samples necessary to provide effective screening for different GM concentrations at the 95% and 99% confidence levels.



Sample sizes

**Calculate Water Volume for Corn Extraction**

Sub-sample weight (g) x 1.5 = volume of water (mL)



Avoid pulling up particles when drawing sample

**Table 1 – Corn: Number of 200 kernel sub-samples required**

Confidence Level	Screening Level % YieldGard Rootworm Corn			
	5%	1%	0.5%	0.25%
95%	1	2	3	6
99%	2	3	5	9

**NOTE:** Screening corn at a 0.5% YieldGard Rootworm concentration level, with 95% confidence, would require testing 3 sub-samples of 200 kernels with all sub-samples negative.

For other sampling scenarios or different screening or confidence levels, refer to the USDA/GIPSA Excel spreadsheet described under Step 1 above, or call EnviroLogix Technical Support for assistance.

**Step 2: Determine Sub-sample Weight, Jar Size, Grind Times and Water Volume Needed for Sample Preparation**

1. Determine the **average weight** of the grain from the lot to be tested. Count and weigh 100 kernels/seeds, then divide by 100.
2. Calculate the sub-sample weight (g) needed for testing, (200 seeds X **average seed weight**).
3. Choose an appropriate jar size for your sample based upon Table 2.
4. Calculate water volume needed for sample preparation. The water volume to sample weight is a ratio of **1.5 to 1**.

*Example Calculation using a 200 kernel sub-sample with an average kernel weight of 0.3g.  
0.3g x 200 = 60 g x 1.5ml = 90 ml water for extraction*

**Table 2**

Commodity	Sub-sample Weight (g)	Jar Size (oz.)	Grind Time (sec.)
Corn	10-25	4	30
	25-65	8	30
	65-250	32	45

**Step 3: Prepare the Sample**

1. Weigh sample into the appropriate size glass Mason jar and attach jar adapter with blade.
2. Place unit on the Waring blender (or equivalent) and cover with protective cover.
3. Grind sample on high speed for specified grinding time or until all whole kernels are broken.
4. Add the volume of tap water calculated by the formula (above left).
5. Cap and shake jar vigorously until the entire sample is wet (20-30 seconds, depending on the number of grains). Sample will begin to settle immediately and liquid can be drawn off at that time.
6. Draw up enough liquid portions from above the settled sample to fill the long narrow tip of the transfer pipette and up to the line at the top of the flared portion of the pipette bulb (see figure at left). Avoid pulling up particles. Dispense extract into reaction vial.
7. To prevent cross-contamination, thoroughly clean blender parts and jars to remove dust and residue prior to preparation of each sample. Use a new transfer pipette and reaction vial for each sample.



## How to Run the QuickStix Strip Test

1. Remove the QuickStix Strip from the canister and reseal the canister. Insert the strip into the sample extract with the arrow tape pointing into the reaction vial.
2. After inserting the strip into the reaction vial, liquid will travel up the membrane strip toward the absorbent pad at the top of the strip. Soon after complete wetting of the membrane, a line will appear on the membrane approximately 1/4 inch below the top absorbent pad. This is the Control Line which should develop within 2-5 minutes. A positive result can be interpreted as soon as a Test Line develops; generally within 2-3 minutes. Allow the strip to develop for the full 5 minutes before concluding that the sample has tested negative.
3. To retain the strip, cut off the bottom section of the stick covered by the arrow tape.

**NOTE:** Reaction vials will stand alone or may be inserted into the cardboard rack provided in the shipment.

## Interpreting the Results

Development of the Control Line within 5 minutes indicates that the strip has functioned properly. Any strip that does not develop a Control Line should be discarded and the sample re-tested using another strip.

If the extract is from a sample containing at least 0.5% YieldGard Rootworm corn (1 kernel in 200), a second line (Test Line) will develop on the membrane strip between the Control Line and the protective tape. *The results should be interpreted as positive for Cry3Bb protein expression.*

If the extract is from a sample containing less than 0.5% YieldGard Rootworm corn, the strip will only develop a control line.



*Any clearly discernable pink Test Line is positive*

## Kit Storage

This QuickStix kit should be stored at room temperature, or refrigerated for longer shelf life. Please note the shelf life on the kit box for each storage temperature. The kit may be used in field applications; however, prolonged exposure to high temperatures may adversely affect the test results.

**Important:** Do not open the desiccated canister until you are ready to use the strips. Allow refrigerated canisters to come to room temperature before opening to prevent condensation. Immediately re-seal unused QuickStix strips in the canister.

## Cross-reactivity

The following materials have been tested with this kit and have been found to cause no false positive results:

- Proteins: Cry1Ab, Cry1Ac, Cry1C, Cry1F, Cry2Aa, Cry2Ab, CP4 EPSPS, Cry9C, PAT/*pat*, PAT/*bar*, and cyanamide hydratase at concentrations at least 4 times higher than reported in commercial crops.



- Other GMO Event Corns: YieldGard® Corn Borer (Bt11, MON810), Knockout®/NatureGard® (Bt176), LibertyLink® (T25), Starlink®, Roundup Ready (Event GA21), Roundup Ready (Event 603), and Herculex™ I corn bulk grain.
- Other Crops: Bulk grain from conventional varieties of corn (yellow or white) and sorghum at 100% of sample; barley, canola, rice, and sugarbeets at 50% of sample; soybean and wheat at 25% of sample; and cotton at 10% of sample.

## Precautions and Limitations

- This kit is designed to screen for presence or absence only, and is not meant to be quantitative.
- This product is currently not applicable for use in any other crop.
- As with all tests, it is recommended that results be confirmed by an alternate method if necessary.
- The assay has been optimized to be used with the protocol provided in the kit. Deviation from this protocol may invalidate the results of the test.
- The results generated through the proper use of this diagnostic tool reflect the condition of the working sample directly tested. Extrapolation as to the condition of the originating lot, from which the working sample was derived, should be based on sound sampling procedures and statistical calculations which address random sampling effects, non-random seed lot sampling effects and assay system uncertainty. A negative result obtained when properly testing the working sample does not necessarily mean the originating lot is entirely negative for the analyte or protein in question.
- Warning: a strong positive result may be safely interpreted in as little as 2 minutes after sample addition. It is not safe, however, to interpret negative results prior to 5 minutes.
- Protect all components from hot or cold extremes of temperature when not in use. Do not leave in direct sunlight or in vehicles.
- Use extreme caution to prevent sample-to-sample cross-contamination with fluids or disposables.



**For Technical Support  
Contact Us At:**

**Envilab GmbH**  
Cyriakusstraße 27


47839 Krefeld

**Tel: (02151) 5671549**

**Fax: (02151) 5695602**

e-mail:  
info@envilab-gmbh.de

website:  
www.envilab-gmbh.de



## LIMITED WARRANTY

EnviroLogix Inc. (“EnviroLogix”) warrants the products sold hereunder (“the Products”) against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product’s printed expiration date. If the Products do not conform to this Limited Warranty and the customer notifies EnviroLogix in writing of such defects during the warranty period, including an offer by the customer to return the Products to EnviroLogix for evaluation, EnviroLogix will repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

**NEITHER ENVIROLOGIX NOR MONSANTO MAKES ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** The warranty provided herein and the data, specifications and descriptions of EnviroLogix products appearing in EnviroLogix published catalogues and product literature are EnviroLogix’ sole representations concerning the Products and warranty. No other statements or representations, written or oral, by EnviroLogix’ employees, agents or representatives, except written statements signed by a duly authorized officer of EnviroLogix Inc., are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

EnviroLogix does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by EnviroLogix, or against damages resulting from such non-EnviroLogix made products or components. EnviroLogix passes on to customer the warranty it received (if any) from the maker thereof of such non-EnviroLogix made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by EnviroLogix.

**THIS WARRANTY IS EXCLUSIVE.** The sole and exclusive obligation of EnviroLogix shall be to repair or replace the defective Products in the manner and for the period provided above. EnviroLogix shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall EnviroLogix be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of EnviroLogix with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

*Roundup Ready and YieldGard are registered trademarks of Monsanto Technology, LLC*

*Knockout is a registered trademark of Novartis AG*

*NatureGard is a registered trademark of Mycogen Corporation*

*LibertyLink is a trademark of Bayer*

*Starlink is a trademark of Aventis Crop Science*

*Herculex is a trademark of Dow AgroSciences LLC*

*EnviroLogix, the EnviroLogix logo, QuickStix and QuickComb are trademarks of EnviroLogix Inc.*

## License

EnviroLogix has developed this kit using proprietary reagents as well as reagents licensed from Monsanto Company.

© EnviroLogix 2006